



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/681,906	10/10/2003	Victor Anisimov	31764	7696

7590
Martin D. Moynihan
PRTSI, Inc.
P.O. Box 16446
Arlington, VA 22215

12/14/2006

EXAMINER

JAWORSKI, FRANCIS J

ART UNIT	PAPER NUMBER
----------	--------------

3768

DATE MAILED: 12/14/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary	Application No. 10/681,906	Applicant(s) ANISIMOV ET AL.	
	Examiner Jaworski Francis J.	Art Unit 3768	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 08 September 2006.
- 2a) ☐ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1 - 15 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☒ Claim(s) 1 - 11 is/are allowed.
- 6) ☒ Claim(s) 12 - 15 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|---|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date <u>8/2/06</u> | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Allowable Subject Matter

Claims 1 – 11 are allowed.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 12 – 15 are rejected under 35 U.S.C. 102(b) as being anticipated by Sugiyama. Sugiyama (US4637256) is directed to three-dimensional tooth scanning using a stepper motor arrangement with a coupled linear array, where in association with the array is a fluid containment housing with membranes 13, 11 adapted to contain a water couplant and be inserted into the mouth. Scanning arm 10 for example may be considered to be that portion of the containment structure which attaches the transducer to the coupling containment.

Lemchen et al (US5278756) as incorporated into Lemchen (US6081739) teaches an ultrasound system in association with a 3D ascension position locator however the

Art Unit: 3768

system as per elements 13, 14 of figs. 2 and 3 of the former use the ultrasound probe as a sonic transponder to effect the locator function.

Tingley et al (US6589054) is directed to the use of curvilinear arrays for localized 2D or 3D tooth scanning without position locator referencing, see cols. 8, 10 mid-portions.

Ylander et al (US5115813) is directed to a bladder (42) coupled ultrasound linear array 36 dental scanner.


Asch et al (US6638219) and Klaassen et al (US6050821) are directed to orthodontic ultrasound mapping systems.

This action is NOT made final however the case should be prepared for final action.

Any inquiry concerning this communication should be directed to Jaworski Francis J. at telephone number 571-272-4738.

FJJ:fjj

12092006


Francis J. Jaworski
Primary Examiner